

Protocol for Neutralizing Antibody Screening Assay for HIV-1 in TZM-bl Cells

(April 2017)

I. Introduction

This assay measures percent neutralization in TZM-bl cells as a function of a reduction in Tat-induced luciferase (Luc) reporter gene expression in the presence of post-immune samples relative to corresponding pre-immune samples after a single round of virus infection. TZM-bl cells (also called JC57BL-13) may be obtained from the NIH AIDS Reagent Program. This is a HeLa cell clone that was engineered to express CD4 and CCR5 [1] and contains integrated reporter genes for firefly luciferase and *E. coli* β -galactosidase under control of an HIV-1 LTR [2], permitting sensitive and accurate measurements of infection. The cells are highly permissive to infection by most strains of HIV, SIV and SHIV, including primary HIV-1 isolates and molecularly cloned Env-pseudotyped viruses. DEAE-Dextran is used in the medium during neutralization assays to enhance infectivity. Expression of the reporter genes is induced in trans by viral Tat protein soon after infection. Luciferase activity is quantified by luminescence and is directly proportional to the number of infectious virus particles present in the initial inoculum. The assay is performed in 96-well culture plates for high throughput capacity. Use of a clonal cell population provides enhanced precision and uniformity. The assay has been validated for single-round infection with either uncloned viruses grown in human lymphocytes or molecularly cloned Env-pseudotyped viruses produced by transfection in 293T/17 cells.

Vaccine-elicited neutralizing antibody responses against tier II viruses tend to be very weak and detection can be missed despite the fact that these low titers could be important. For maximum sensitivity and specificity, post-immune samples can be screened at a single dilution, in triplicate, to measure the reduction in RLU relative to wells that contain the same dilution of the corresponding pre-immune samples. This way, neutralization is not measured relative to the virus control wells as is the case for the Neutralizing Antibody Assay for HIV-1 in TZM-bl cells where samples are assayed at multiple dilutions. Use of corresponding pre- and post-immune samples in the screening assay gives a more precise measure of true neutralization because it adjusts for non-specific effects. For example, values <50% neutralization in the screening assay may be significant if they exceed the positive deflection which may be seen in the placebo group.

II. Definitions

GM: Growth Medium

Luc: Luciferase

RLU: Relative Luminescence Units

DPBS: Dulbecco's Phosphate Buffered Saline

ID: Identification

TCID: Tissue Culture Infectious Dose

DEAE-Dextran: Diethylaminoethyl-Dextran

EDTA: Ethylenediaminetetraacetic acid

III. Reagents and Materials

Recommended vendors are listed. Unless otherwise specified, products of equal or better quality than the recommended ones can be used whenever necessary.

TZM-bl Cells

Vendor: NIH AIDS Research and Reference Reagent Program

Growth Medium (see Protocol for Reagent Preparation for Use in the Neutralizing Antibody Assay for HIV-1 in TZM-bl Cells)

DEAE-Dextran, hydrochloride, average Mol. Wt. 500,000 (see Protocol for Reagent Preparation for Use in the Neutralizing Antibody Assay for HIV-1 in TZM-bl Cells)

Vendor: Sigma

Trypsin-EDTA (0.25% trypsin, 1 mM EDTA) (see Protocol for Trypsin-EDTA Treatment for Disruption of Cell Monolayers)

Vendor: Invitrogen

Sterile

Trypan Blue (0.4%)

Vendor: Sigma

DPBS

Vendor: Invitrogen

Sterile

Britelite Plus Reporter Gene Assay System (see Protocol for Reagent Preparation for Use in the Neutralizing Antibody Assay for HIV-1 in TZM-bl Cells)

Vendor: Perkin Elmer Life and Analytical Sciences

NOTE 1: Bright Glo substrate solution from Promega and Britelite substrate solution from Perkin Elmer Life and Analytical Sciences are acceptable substitutes for Britelite Plus. Please follow manufacturer's guidelines for preparation and use. Britelite and Bright Glo are classified as hazardous. Personal Protective Equipment (PPE) is required when working with these reagents.

Microliter pipettor tips, sterile

Vendor: VWR

Vendor: Rainin

Vendor: Fisher

Disposable pipettes, sterile, individually wrapped

Vendor: Fisher

1 ml pipettes

2 ml pipettes

5 ml pipettes

10 ml pipettes

25 ml pipettes

50 ml pipettes
100 ml pipettes

Flat-bottom culture plates, 96-well, low evaporation, sterile

Vendor: Fisher

Flat-bottom black solid plates, 96-well

Vendor: Costar/Fisher

Culture flasks with vented caps, sterile

Vendor: Fisher

T-25 flask

T-75 flask

Reagent reservoirs, 50 ml, 100 ml capacity

Vendor: Costar

Vendor: VWR

IV. Instrumentation

Recommended manufacturers are listed. Unless otherwise specified, equipment of equal or better quality than the recommended ones can be used whenever necessary.

Biological Safety Cabinet

Manufacturer: Baker Co.

Incubator

Manufacturer: Forma Scientific

Water-jacketed (37°C, 5% CO₂ standard requirements)

Centrifuge and Microcentrifuge

Manufacturer: Jouan

(low speed capable of up to 500 x g)

50 ml tube holder

15 ml tube holder

Microtitration plate holder

Manufacturer: Eppendorf

18 place standard rotor F-45-18-11 for 1.5 ml microcentrifuge tubes

Luminometer

Manufacturer: PerkinElmer Life Sciences

Water bath

Manufacturer: Precision Scientific

Hemocytometer

Manufacturer: INCYTO

NOTE 2: An automated cell counting device (i.e.. Countess, Manufacturer: Invitrogen) may be used in lieu of a light microscope / hemocytometer for cell counting and viability calculation.

Pipettor

Manufacturer: ThermoLabsystem

12-channel pipette, 5-50 μ l

12-channel pipette, 30-300 μ l

Single channel pipette, 5-50 μ l

Single channel pipette, 30-200 μ l

Manufacturer: Drummond Scientific Co.

PipetteAid XP

Manufacturer: BioHit

12 channel, 50-1200 μ l Electronic Pipette

Single channel, 10-300 μ l Electronic Pipette

Single channel, 5-120 μ l

Manufacturer: Rainin

12 channel pipettor, 20-200 μ l

Light Microscope

Manufacturer: Olympus

Low Temperature Freezer (-70°C or lower)

Manufacturer: Harris

Manufacturer: Puffer Hubbard

4 $^{\circ}$ C Refrigerator

Manufacturer: Sci-Cool

-20°C Freezer

Manufacturer: Sci-Cool

V. Specimens

Samples should be heat-inactivated at 56°C as described in Protocol for Heat-Inactivation of Serum and Plasma Samples. Samples may be serum or plasma, although serum is preferred. Anticoagulants in plasma are problematic in the assay, especially when heparin is used. For example, some forms of heparin have potent and strain-specific antiviral activity. Also, all anticoagulants can be toxic to the cells at low plasma dilutions.

VI. Protocol

1. Neutralization Assay

NOTE 3: All incubations are performed in a humidified 37°C , 5% CO_2 incubator unless otherwise specified.

NOTE 4: The exterior-most wells in rows A and H of columns 3-12 are not utilized for the assay but serve as blank wells. The entire column 12 also serves as blank wells (see Appendix A)

1.1 Using the format of a 96-well flat bottom culture plate as illustrated in Appendix A, place 150 μ l of GM in all wells of column 1 (cell control). Place 100 μ l in all wells of column 2 (column 2 will be the virus control). Place an additional 85 μ l in all wells of columns 3-11, rows B-G, (to receive test samples). Place 200-250 μ l in rows A and H of columns 3-11. Place 200-250 μ l in column 12 to serve as the blank.

1.2 Centrifuge test samples to pellet any debris that might be present..

1.3 Add 15 μ l of test sample in triplicate beginning in Row G, according to the plate layout described in Appendix A.

NOTE 5: The above description is for a dilution of 1:10. Appropriate adjustments may be made to test a different dilution.

1.4 A positive control with a known neutralization titer against the target virus should be assayed on at least one plate in series each time assays are performed using the method described in Protocol for Neutralizing Antibody Assay for HIV-1 in TZM-bl Cells.

1.5 Thaw the required number of vials of virus by placing in an ambient temperature water bath placed in a biological safety cabinet. When completely thawed, dilute the virus in GM to achieve a TCID of approximately 150,000 RLU equivalents (+/- 15,000 RLU) per 50 μ l. For pseudoviruses that do not reach 150,000 RLU, pick a dose of virus that gives at least 15,000 RLU but is not toxic to the cells via light microscopy. See Protocol for Preparation and Titration of HIV-1 Env-pseudotyped Viruses for measurement of TCID in TZM-bl cells.

NOTE 6: The RLU equivalents measured in the TCID assay will not necessarily match the RLUs obtained in the virus control of the neutralization plate. This difference is acceptable provided that the virus control is $\geq 10X$ the background in cell control wells and the virus control is not toxic to the cells based on light microscopy.

NOTE 7: Leftover virus may be refrozen and stored at -80°C but first should be marked with a "1X" on the lid and label of the vial. The "1X" notes that that particular vial has been thawed one time. When using "1X" vials of virus in the TZM-bl assay, the technician must consult the virus database to obtain the optimal virus dilution for viruses that have been thawed one time. No pseudovirus should be used in the TZM-bl assay if it has been thawed and refrozen more than once.

1.6 Dispense 50 μ l of cell-free virus to all wells in columns 2-11, excluding wells in rows A and H in columns 3-11.

Virus Calculations:

To calculate the total volume (vol.) of virus/GM mixture needed for the assay, multiply the total number of plates by the volume of virus/GM mixture to be used per plate. Then divide the total volume of virus/GM mixture by the optimal virus dilution to use (based on the TCID assay) to derive the volume of undiluted virus needed. Then subtract the volume of undiluted virus needed from the total volume of virus/GM mixture to derive the volume of GM needed.

Total number of plates X Vol. of virus/GM per plate = Total vol. virus/GM needed

Total vol. virus/GM needed ÷ Optimal virus dilution = Vol. of undiluted virus needed

Total vol. of virus/GM needed – Vol. of undiluted virus needed = Vol. of GM needed.

1.7 Cover the plates and incubate for 45 – 90 minutes.

1.8 During the incubation, trypsinize the cells 10-20 minutes prior to use as described in Protocol for Trypsin-EDTA Treatment for Disruption of Cell Monolayers. Following, prepare a suspension of TZM-bl cells at a concentration of 1×10^5 cells/ml in GM as described below:

1.8.1 Perform viable cell count (use laboratory specific protocol)

1.8.2 Cell Calculations:

To calculate the cell concentration, count the total number of cells in a predetermined number of quadrants in a hemacytometer and obtain the average cell count per quadrant. Multiply this number by the dilution factor to yield the cell concentration, “ C_1 ”, in cells $\times 10^4$. To calculate the total cell mixture volume, “ V_2 ”, that you need, multiply the number of plates by the total volume of cell mixture needed per plate. The concentration of cells desired is 100,000 cells/ml, “ C_2 ”. Thus, using the equation $C_1V_1 = C_2V_2$, one can solve for “ V_1 ”, the volume of cells needed.

For example:

Total number of cells counted = 60

Number of quadrants counted = 4

Dilution factor = 10

Number of plates = 1

Cell mixture needed per plate = 10 ml

$60 \text{ cells} \div 4 \text{ quadrants} = 15 \text{ cells/quadrant}$

$15 \times 10 = 150 \times 10^4 \text{ cells/ml} = 15 \times 10^5 \text{ cells/ml} = C_1$

$1 \text{ plate} \times 10 \text{ ml/plate} = 10 \text{ ml} = V_2$

Optimum final concentration of cells = 100000 = C_2

Therefore: $C_1V_1 = C_2V_2$

$(100000 \times 10) \div 1500000 = 0.67 \text{ ml of cells}$

1.8.3 Addition of DEAE-Dextran to Cells

The DEAE-Dextran concentration in the cell suspension will be 25 µg/ml. The final concentration of DEAE-Dextran in the assay plate will be 10 µg/ml.

To calculate the amount of DEAE-Dextran to use, first multiply the optimal concentration of DEAE-Dextran (see Protocol for Determination of Optimal Concentration of DEAE-Dextran) by 0.250 ml (the final volume in each well) to get the amount of DEAE-Dextran per well. Multiply the amount of DEAE-Dextran per well by 100 wells/plate (96 wells rounds to 100) to derive the amount of DEAE-Dextran per plate. Divide the amount of DEAE-Dextran needed per plate by the stock concentration of the DEAE-Dextran to yield the volume of DEAE-Dextran stock needed. Multiply this number by the number of plates to yield the total volume of DEAE-Dextran stock needed.

For example:

If the optimal concentration of DEAE-Dextran to use is 10 µg/ml and the DEAE-Dextran stock is at 5 mg/ml

$10 \mu\text{g/ml} \times 0.25 \text{ ml (volume in well)} = 2.5 \mu\text{g}$ of DEAE-Dextran needed in each well

$2.5 \mu\text{g} \times 100 \text{ wells/plate} = 250 \mu\text{g}$ of DEAE-Dextran needed per plate = 0.25 mg of DEAE-Dextran

$0.25 \text{ mg of DEAE-Dextran per plate} \div 5 \text{ mg/ml stock concentration} = 0.05 \text{ ml}$ of DEAE-Dextran stock needed per plate

To calculate the amount of Growth Medium to add, subtract the total volume of cells needed and the total volume of DEAE-Dextran stock needed from the total volume of cell mixture needed.

The total volume needed for one plate is 10 ml

$10 \text{ ml} - 0.67 \text{ ml cells} - 0.05 \text{ ml DEAE-Dextran} = 9.28 \text{ ml of GM}$

1.9 The GM/cells/DEAE-Dextran suspension should be prepared as follows: Add GM and DEAE-Dextran and mix. Add cells and thoroughly mix the prepared cell suspension immediately prior to plating. Dispense 100 µl of the prepared cell suspension (10,000 cells per well) to each well in columns 1-11, rows A and H, excluding wells in rows A and H, columns 3-11.

NOTE 8: The concentrations of DEAE-Dextran shown above will vary by batch of DEAE-Dextran. The actual optimal concentration should be determined for each new batch of DEAE-Dextran prepared in accordance with Protocol for Determination of Optimal Concentration of DEAE-Dextran.

NOTE 9: The use of DEAE-Dextran is optional. When omitted, the TCID of the virus should be measured in the absence of DEAE-Dextran.

1.10 Cover plates and incubate for 48 – 72 hours if Env-pseudotyped viruses are used. If replication-competent virus is used, the plates should be incubated for 46-50 hours to minimize virus replication.

1.11 After incubation, remove plates from the incubator. Plates should not stay out of the incubator longer than one hour before running the luciferase reaction.

NOTE 10: Examine at least 2 virus control wells for the presence of syncytia by microscopic examination. It is important to note the presence of syncytia as too many syncytia indicate cell killing and thus the validity of the assay is compromised. If cell killing is present, the assays should be repeated using a lower dose of the virus. It is also important that all wells of the plate containing test sample be checked for the presence of toxicity. Cell toxicity could be erroneously interpreted as neutralization.

1.12 Thaw Britelite Plus directly before use in an ambient temperature water bath away from light.

1.13 Remove 150 µl of culture medium from each well, leaving approximately 100 µl.

1.14 Dispense 100 µl of Britelite Plus Reagent to each well in columns 1-11, excluding wells in rows A and H in columns 3-11.

1.15 Incubate at room temperature for 2 minutes to allow complete cell lysis. Mix by pipettor action (at least two strokes) and transfer 150 µl to a corresponding 96-well black plate. Read the plate after the two minute incubation time (but no longer than fifteen minutes) in a luminometer.

2. Analyzing and printing results

2.1 Prior to reading the plates in the luminometer, enter the assay protocol information in the Wallac Software of the luminometer.

2.2 Read the plates in a luminometer interfaced to a dedicated computer in the laboratory.

2.3 Use the software program associated with the luminometer to save the raw data onto the desired location, after each plate is read, using a unique file identification number (ID) for each plate.

2.4 Analyze and print the data using the modified screening assay macro provided by the Central Reference Laboratory. The data print-out must include: i) experiment number, ii) protocol and/or study number, iii) cells used in the assay, iv) length of incubation, v) name, lot number and dilution of the virus stock used, vi) ID, visit number and bleed date of each sample and vii) signature of technician who performed the assay.

2.5 Percent neutralization is determined by calculating the difference in average RLU between test wells containing post-immune sample and test wells containing pre-immune sample from the same individual. The pre-immune and post-immune samples must be assayed on the same assay plate.

3. Pass/Fail Criteria

3.1 The average RLU of virus control wells is >10 times the average RLU of cell control wells.

3.2 The percent CV (%CV) of RLU in the virus control wells is $\leq 30\%$.

3.3 The percent CV (%CV) for the triplicate wells is $\leq 30\%$ for sample dilutions that yield at least 40% neutralization.

3.4 The value of the positive control is within 3-fold of the average of the previous values for that particular control/virus combination.

NOTE 11: A final decision regarding the acceptability of the data will be made by the P.I. (or his/her designee).

4. References

1. Platt, E.J., K. Wehrly, S.E. Kuhmann, B. Chesebro, and D. Kabat. 1998. Effects of CCR5 and CD4 cell surface concentrations on infection by macrophage tropic isolates of human immunodeficiency virus type 1. *J. Virol.* 72:2855-2864.
2. Wei, X., J.M. Decker, H. Liu, Z. Zhang, R.B. Arani, J.M. Kilby, M.S. Saag, X. Wu, G.M. Shaw, and J.C. Kappes. 2002. Emergence of resistant human immunodeficiency virus type 1 in patients receiving fusion inhibitor (T-20) monotherapy. *Antimicrob. Agents Chemother.* 46:1896-1905.

Appendix

A: Assay template for screening at a single serum dilution

	1	2	3	4	5	6	7	8	9	10	11	12
A	CC	VC	BLK	BLK	BLK	BLK	BLK	BLK	BLK	BLK	BLK	BLK
B	CC	VC	S #3 Post	S #3 Post	S #3 Post	S #6 Post	S #6 Post	S #6 Post	S #9 Post	S #9 Post	S #9 Post	BLK
C	CC	VC	S #3 Pre	S #3 Pre	S #3 Pre	S #6 Pre	S #6 Pre	S #6 Pre	S #9 Pre	S #9 Pre	S #9 Pre	BLK
D	CC	VC	S #2 Post	S #2 Post	S #2 Post	S #5 Post	S #5 Post	S #5 Post	S #8 Post	S #8 Post	S #8 Post	BLK
E	CC	VC	S #2 Pre	S #2 Pre	S #2 Pre	S #5 Pre	S #5 Pre	S #5 Pre	S #8 Pre	S #8 Pre	S #8 Pre	BLK
F	CC	VC	S #1 Post	S #1 Post	S #1 Post	S #4 Post	S #4 Post	S #4 Post	S #7 Post	S #7 Post	S #7 Post	BLK
G	CC	VC	S #1 Pre	S #1 Pre	S #1 Pre	S #4 Pre	S #4 Pre	S #4 Pre	S #7 Pre	S #7 Pre	S #7 Pre	BLK
H	CC	VC	BLK	BLK	BLK	BLK	BLK	BLK	BLK	BLK	BLK	BLK

CC, Cell control wells (cells only). VC, virus control wells (virus and cells but no serum sample are added here). BLK, blank wells. S #, Sample number.

	1	2	3	4	5	6	7	8	9	10	11	12
A	CC	VC	BLK	BLK	BLK	BLK	BLK	BLK	BLK	BLK	BLK	BLK
B	CC	VC	S #2 Post	S #2 Post	S #2 Post	S #4 Post	S #4 Post	S #4 Post	S #6 Post	S #6 Post	S #6 Post	BLK
C	CC	VC	S #2 Post	S #2 Post	S #2 Post	S #4 Post	S #4 Post	S #4 Post	S #6 Post	S #6 Post	S #6 Post	BLK
D	CC	VC	S #2 Pre	S #2 Pre	S #2 Pre	S #4 Pre	S #4 Pre	S #4 Pre	S #6 Pre	S #6 Pre	S #6 Pre	BLK
E	CC	VC	S #1 Post	S #1 Post	S #1 Post	S #3 Post	S #3 Post	S #3 Post	S #5 Post	S #5 Post	S #5 Post	BLK
F	CC	VC	S #1 Post	S #1 Post	S #1 Post	S #3 Post	S #3 Post	S #3 Post	S #5 Post	S #5 Post	S #5 Post	BLK
G	CC	VC	S #1 Pre	S #1 Pre	S #1 Pre	S #3 Pre	S #3 Pre	S #3 Pre	S #5 Pre	S #5 Pre	S #5 Pre	BLK
H	CC	VC	BLK	BLK	BLK	BLK	BLK	BLK	BLK	BLK	BLK	BLK

CC, Cell control wells (cells only). VC, virus control wells (virus and cells but no serum sample are added here). BLK, blank wells. S #, Sample number